

Missouri Delta Community Hospital

Sikeston, Missouri

63801

July 12, 1982

30-02377

J. R. Miller, Chief  
Technical Inspection Branch  
Nuclear Regulatory Commission  
Region III  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

Lic. 24-12876-02

Dear Sir:

The annual accuracy check of the dose calibrator will be done in October which is the yearly anniversary date of NBS source. This was recommended by the Inspector because it would be easier to account for the decay of the Cs-137 source. It will start this October, 1982. (This has been and will continue to be done daily and logged. This log was shown to the Inspector.) The test will be done using both Cs-137 and Co-57 NBS sources every October and kept on file for inspection.

Sincerely,

*Jerry Costin, R.T.*

Jerry Costin, R. T.  
Chief Technologist

JC:mjv

CC: Harold Jones, Administrator

IE HQ FILE COPY