

Leeco Diagnostics, Inc.

DCD/DCB

24475 W. 10 MILE ROAD • SOUTHFIELD, MI 48034
(313) 353-2620 • 1-800-950-2620 • FAX (313) 353-5039

November 6, 1990

License File
21-18294-01

John A. Grobe, Chief
Nuclear Materials Safety Branch
U.S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, IL 60137

Re: NRC Byproduct Material License No. 21-18294-01

Dear Mr. Grobe:

Further to a recent inspection of Leeco Diagnostics (September 26, 1990) conducted by Mr. James L. Lynch and your Notice of Violation (October 26, 1990), the following corrective actions to the noted violations have been instituted.

LICENSE CONDITION NO. 14

Violation 1

- Both film badges and TLD ring badges are being worn by the Hot Lab employees at all times when handling radioactive material (refer to Attachment No. 1).

Violation 2

- In lieu of an amendment to the license, a letter was addressed to Ms. Toye L. Simmons on July 3, 1989 (Attachment No. 2) explaining/justifying the discontinuation of the activated charcoal filters. However, a formal amendment was requested on October 15, 1990 (Attachment No. 3) for discontinuance of activated charcoal filters in the Hot Lab exhaust hoods.

Effluent data in both instances indicate that the effluent is well below maximum permissible concentrations for unrestricted areas.

As a point of clarification the undersigned has a direct reporting responsibility to the CEO for all Regulatory functions and a dotted line responsibility for Quality Assurance functions (refer to Attachment No. 4)

9012030095 901106
REG3 LIC30
21-18294-01
FDC


IED 711

NOV 8 1990

John A. Groka
November 6, 1990
Page 2

If you require any further explanation please do not hesitate to contact me at 313-353-2620.


I remain,



Dante J. Capaldi, Ph.D.
Director, Quality Assurance/Regulatory Affairs

DJC/mbc

cc: NRC File
Rod Raynovich



ATTACHMENT #1

Leers Diagnostics, Inc.

24475 W. 10 MILE ROAD • SOUTHFIELD, MI 48034
(313) 353-2620 • 1-800-950-2620 • FAX (313) 353-5039

October 1, 1990

MEMORANDUM

To: Hot Lab Employees
From: Dante J. Capaldi ✓
Subject: Film Badges and TLD Ring Badges

Please be reminded that film badges and TLD ring badges are to be worn at all times when working with radioactive material.

Furthermore, eye protection and at least two pairs of disposable latex gloves are to be worn during the iodination procedure.

Please post this notice on your bulletin board or the exhaust hood. This will serve as a constant reminder of a very important precaution.

DJC/mbc

cc: NRC File
Rod Raynovich
Ron Steckel
Joyce Chang