DEPARTMENT OF RADIOLOGY ABINGTON MEMORIAL HOSPITAL Abington, Pennsylvania 19001 215-885-3110

DOCKETED USNRC

AGO -9 P1

August 6, 1982

OFFICE OF BRANC

Secretary of the Commission J.S. Nuclear Regulatory Commission Washington D.C. 20555

PROPOSED RULE PR-Misc. No fice (Reg Guide)

ATTENTION: JOCKETING AND SERVICE BRANCH

Commentary on the ) raft Regulatory Guide for License in Medical Tele-Therapy Programs.

Item 13C "Calibration of Dosime rv Systems".

The calibration of the dosimetry systems at two year interval is impractical for several reasons:

- The NBS and the few Regional Calibration Laboratories a). acredited by the AAPM are overloaded with instruments to be calibrated. A waiting period of six months to on year to have the system accepted for calibration is not uncommon.
- Institutions with only one dosimetry system can be b) several months without any instructions at all. The reason for that is the time required for the chambers and electrometers to be at the laboratory before calibration and the delay in receiving the writing report after the calibration is done, indicating the proper correcting factors.

Because the monthly calibration of the CO-60 unit serves as the periodical calibration check of the instructment, a four year interval between NBS or regional certified calibration laboratory check of the dosimetry systems will be more practical.

Certified Medical Physicist

mp

RADIATION THERAPY AND NUCLEAR MEDICINE: Emanuel G. Tuisky, M.D. Alan Cohler, M.D.

MEDICAL PHYSICIST: Victoria G. Castro, D.E. M.S. John M. Niemkiewicz, M.S.

8208160293 820806 PDR REGGD PDR

DIAGNOSIS:

Charles H. Sillars, M.D. John W. Breckenridge, M.D. Mark D. Olken, M.D. Ruth E. Buck, M.D. Stephen Pripstein, M.D. F. E. Hanby, M.D. John B. Harper, M.D.

Leopoldo E. Margarida, M.D. Jere F. Seelaus, M.D. Paul A. Urffer, M.D.

Acknowledged by card. S.