

DCD/DCB

NRC FORM 591 PART 1
(7-91)
10 CFR 2.201

U.S. NUCLEAR REGULATORY COMMISSION

SAFETY INSPECTION

Page 1 of 1

1. LICENSEE 2222 Medical Imaging 6-2222 South Linden Road Flint, MI 48532		2. REGIONAL OFFICE REGION III U S NUCLEAR REGULATORY COMMISSION 799 ROOSEVELT ROAD GLEN ELLYN IL 60137	
3. DOCKET NUMBER(S) 030-30606	4. LICENSE NUMBER(S) 21-25921-01	5. DATE OF INSPECTION November 23, 1993	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- 1. Within the scope of this inspection, no violations were observed.
 - 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.
 - 3. During this inspection certain of your activities, as described below or attached, were in violation of NRC requirements. This form is a **NOTICE OF VIOLATION**, which is required to be posted in accordance with 10 CFR 19.11.
 - A. _____ was not properly posted to indicate the presence of a _____ 10 CFR 20.203(b),(c),(d),(e) or 34.42.
 - B. _____ of sealed sources were not performed at the proper frequencies. 10 CFR _____ or License Condition Number _____.
 - C. Records of Decay-in-storage were not properly maintained. 10 CFR 35.92 (B) or License Condition Number _____.
 - D. Documents were not properly posted or otherwise made available. 10 CFR 19.11.
 - E. Reports or notification of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____.
- Since January 1993, licensee failed to record background dose rates and dose rates measured at surface of radioactive waste containers held for decay-in-storage prior to disposal.

IR-07
PUBLIC

200641
9312210096 931123
PDR ADOCK 03030606
C PDR

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

SIGNATURE -- LICENSEE <i>R. Weigand</i>	DATE 11/23/93	SIGNATURE -- NRC INSPECTOR <i>J. H. O'Kitt</i>	DATE 11/23/93
--	------------------	---	------------------