

DCD/DCB

DEC 08 1993

LICENSE FILE

Bronson Methodist Hospital
ATTN: Patrick Ludwig
President and CEO
252 East Lovell Street
Kalamazoo, MI 49007

License No. 21-13125-01
Docket No. 030-02146

Dear Mr. Ludwig:

This refers to our review of your revised Quality Management Program (QMP) for Radiopharmaceutical and Brachytherapy use dated August 1993. That revision was submitted in response to the NRC evaluation of your January 1992 QMP provided to you on July 22, 1993.

Our review of your revised QMP was performed to determine whether policies and procedures have been developed to meet the objectives of 10 CFR 35.32. Based on your submissions, your QMPs of January 1992 and August 1993, together, appear to meet the requirements of the rule. Please combine the January 1992 QMP and the August 1993 QMP to create one complete document.

No reply is required in response to this letter. We will review implementation of your QMP at our next regular inspection of your facility.

Please be advised that this QMP will not be incorporated into our license through the "tie-down" condition. This allows you the flexibility to make changes to your quality management program without obtaining prior NRC approval. When modifications are made to your QMP, you should submit those changes to this Office within 30 days, as required by 10 CFR 35.32(e).

Thank you for your cooperation in this matter. If you have any questions, please call me at (708) 790-5721.

Sincerely,

John A. Grobe, Chief
Nuclear Materials Inspection
Section 2

200043

cc: Clarence Snyder, Radiology Director
Mark Watts, Radiation Safety Office

bcc: S. Merchant, NMSS
PUBLIC

RIII
EMK
Matson/jaw
12-6-93

RIII
[Signature]
Frazier
12-6-93

RIII
[Signature]
Grobe
12-6-93

IE07
110
PUBLIC

9312210119 931203
PDR ADOCK 03002146
C PDR