

LICENSEE EVENT REPORT

CONTROL BLOCK: _____ (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

01 I L Z I S 1 00 - 0 0 0 0 0 - 0 0 3 4 1 1 1 1 4 5
7 8 9 LICENSEE CODE 14 15 LICENSE NUMBER 25 26 LICENSE TYPE 30 57 CAT 58

CON'T

01 REPORT SOURCE L 6 0 5 0 0 0 2 9 5 1 1 1 7 8 1 8 1 2 0 8 8 1 9
7 8 9 DOCKET NUMBER 68 69 EVENT DATE 74 75 REPORT DATE 80

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (3)

02 While performing daily rad. monitor check source testing (PT-0, Appx. N), gas decay tank particulate rad. monitor ORT-PR10C would not respond to its check source or to an external source. The monitor was taken O.O.S., and shiftly particulate samples were taken as required by T.S. 4.12.1.A. No gas decay tank release was in progress. There were no safety implications, and the health and safety of the public were not affected.

09 SYSTEM CODE M C 11 CAUSE CODE E 12 CAUSE SUBCODE G 13 COMPONENT CODE I N S T R U 14 COMP. SUBCODE P 15 VALVE SUBCODE Z 16
7 8 9 10 11 12 13 14 15 16

17 LER/RO REPORT NUMBER 8 1 21 22 EVENT YEAR 8 1 23 SHUTDOWN METHOD 24 SEQUENTIAL REPORT NO. 0 4 7 26 OCCURRENCE CODE 27 ATTACHMENT SUBMITTED 0 3 28 29 REPORT TYPE L 30 REVISION NO. 0 32

ACTION TAKEN FUTURE ACTION EFFECT ON PLANT SHUTDOWN METHOD HOURS ATTACHMENT SUBMITTED NPD-4 FORM SUB. PRIME COMP. SUPPLIER COMPONENT MANUFACTURER
E 18 Z 19 Z 20 Z 21 0 0 0 0 22 N 23 N 24 Z 25 N 3 0 5 26
33 34 35 36 37 38 39 40 41 42 43 44 45 46 47

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

10 The high voltage supply to the rad. monitor detector was found low by 300 V. It is unknown precisely how this occurred. The voltage was readjusted, and the monitor recalibrated and returned to service. This is the first occurrence of this type. No further action is necessary.

15 FACILITY STATUS E 28 % POWER 1 0 0 29 OTHER STATUS NA 30 METHOD OF DISCOVERY B 31 DISCOVERY DESCRIPTION Operator testing 32
7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 27 28 29 30 31 32

16 ACTIVITY CONTENT Z 33 Z 34 AMOUNT OF ACTIVITY NA 35 LOCATION OF RELEASE NA 36
7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 27 28 29 30 31 32

17 PERSONNEL EXPOSURES NUMBER 0 0 0 37 TYPE Z 38 DESCRIPTION NA 39
7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 27 28 29 30 31 32

18 PERSONNEL INJURIES NUMBER 0 0 0 40 DESCRIPTION NA 41
7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 27 28 29 30 31 32

19 LOSS OF OR DAMAGE TO FACILITY TYPE Z 42 DESCRIPTION NA 43
7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 27 28 29 30 31 32

20 PUBLICITY ISSUED DESCRIPTION N 44 NA 45
7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 27 28 29 30 31 32

NAME OF PREPARER Fred Ost PHONE: 312-746-2084, Ext 555

8112170095 811208
PDR ADOCK 05000295
S PDR

NRC USE ONLY