

...the close  
advantage



TAHLEQUAH  
CITY HOSPITAL

December 27, 1993

U.S.N.R.C., Region IV  
Attn: Document Control Desk  
Washington, D.C. 20555

Reply to a Notice of Violation  
License #35-15626-01  
Docket #30-09481/93-01

The following information is in response to the NRC correspondence of December 15, 1993. Each reply listed below corresponds to the NRC correspondence enumeration code.

- A.
1. This violation is acknowledged. The Nuclear Medicine technologist is registered with more than fifteen years experience as a nuclear technologist. She says she misunderstood the requirements of the weekly area wipe test. The physicist reviewed the documentation quarterly.
  2. The physicist discussed this violation with the nuclear technologist. Both agree that she is properly making the weekly wipe tests at this time.
  3. The physicist will be giving a comprehensive training program to all nuclear personnel on "routine" calibrations, surveys, assays, and radioisotope controls on January 19. The training will be documented and filed.
  4. We believe this violation has been corrected as of this date.
- B.
1. This violation is acknowledged. A 3" X 5" card stating that "Title 10 parts 19,20,30 and 35 are stored in the right, bottom drawer of the nuclear medicine desk with a copy of the nuclear license and

000001  
9401070355 931227  
PDR ADDCK 03009481  
C PDR

IE07  
1/0

the license application" has been posted on the nuclear medicine bulletin board for at least eight years. We are unable to explain why it is missing at this time.

2. As of this date, a new card stating the same information has been prepared and displayed on the nuclear medicine bulletin board.
3. The proper display of this card, the license, and other required documents will be checked each quarter during the quarterly nuclear medicine record keeping audit. Assurance that 10 CFR parts 19, 20, 30 and 35 are available will also be made during those audits.
4. This violation has been corrected as of this date.

If you need further action or information concerning this matter, please contact me at (918) 456-0641.

Sincerely,



L. Gene Matthews  
Administrator