

Appendix

NOTICE OF VIOLATION

Wisconsin Power & Light Company

License No. 48-18772-02

As a result of the routine safety inspection conducted on August 11, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified:

1. License Condition 10 states that licensed material shall be used only at 814 South Division Street, Waunakee, WI.

Contrary to the above, licensed material has been relocated to 800-D South Division Street, Waunakee, WI, a location not authorized on your license.

This is a Severity Level V violation (Supplement VI).

2. License Condition 14 requires that each chromatograph detector cell containing nickel-63 be tested for leakage and/or contamination at intervals not to exceed six (6) months.


Contrary to the above, neither your Tracor 560 gas chromatograph (License No. 48-18772-02) nor your Varian 6000 gas chromatograph (General License) have been tested for leakage and/or contamination at six (6) month intervals.

This is a Severity Level V violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

8-16-82

Dated



D. G. Wiedeman, Chief
Materials Radiation Protection
Section 1