

Appendix

NOTICE OF VIOLATION

Unity Hospital

License No. 21-18556-01

As a result of the inspection conducted on November 23, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violations were identified:

1. License Condition No. 15 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

The referenced application dated June 29, 1978, states in Item 10 that the dose calibrator will be checked for constancy each day of use.

Contrary to the above, the dose calibrator was used on at least nine occasions between July 1981, and September 1981, and was not tested for constancy of operation. This is a repeat item of noncompliance.

This is a Severity Level V violation (Supplement VII).

2. License Condition No. 15 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

The referenced application dated June 29, 1978, states in Item 17 that all laboratory areas will be surveyed weekly and this survey will consist of a G-M survey and a series of wipe tests.


Contrary to the above, a wipe test of laboratory areas in use have not been performed since March 23, 1981.

This is a Severity Level V violation (Supplement VII).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation. Consideration may be given to extending your response time for good cause shown.

Dated

12/10/81


D. J. Sreniawski, Chief
Materials Radiation Protection
Section 2