

CBP

AUG 18 1982

Docket No.: 30-04639

License No.: 20-04490-01

Center for Blood Research  
ATTN: Douglas Surgenor  
Director  
800 Huntington Avenue  
Boston, Massachusetts 02115

Gentlemen:

Subject: Inspection No. 30-04639

This refers to your letter dated July 12, 1982, in response to our letter dated June 11, 1982.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

From a July 20, 1982 telephone conversation between Dr. L. F. Friedman of this office and Dr. F. J. Lionetti of your staff we understand that Dr. Lionetti will institute procedures to assure that survey instruments are calibrated on schedule and that wipe tests of incoming packages will be performed. If this is not in accordance with your understanding, please contact this office.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:  
John D. Kinneman

*for* Thomas T. Martin, Director  
Division of Engineering and Technical  
Programs

cc:  
Public Document Room (PDR)  
Nuclear Safety Information Center (NSIC)  
Commonwealth of Massachusetts (2)  
Dr. F. Lionetti, Senior Investigator and Radiation Safety Officer

bcc:  
Region I Docket Room (with concurrences)

*for* *Kinneman*  
RI:DETP  
Johansen/ntm  
8/06/82

*Kinneman*  
RI:DETP  
Kinneman

*8/16/82*

OFFICIAL RECORD COPY

8208230269 820818  
NMS LIC30  
20-04490-01 PDR

1E:07