



Center for Blood Research Inc.

800 Huntington Ave., Boston, MA 02115 · (617) 731-6470

July 12, 1982

Mr. Thomas T. Martin
Director
Division of Engineering and Technical Programs
United States Nuclear Regulatory Commission
Region 1
631 Park Avenue
King of Prussia, Pennsylvania 19406

Dear Mr. Martin:

This is to reply to your letter of June 11, 1982 and Notice of Violation (Docket 30-04639 License number 20-04490-01). The following action has been taken.

I have met with (our consultant) Robert Johnson of the Radiological Services of Harvard University to derive his advice and assistance.

We will henceforth maintain a running inventory of radioactive materials on hand. There will be one person who will authorize purchase of isotopes. Copies of purchase orders will be on file and also sent to Mr. Johnson.

A quarterly audit of our radiation safety program will be conducted here by Mr. Johnson.

With respect to Condition 14 of License 20-04490-01:

Item 11: Calibration of survey instruments -- This has been arranged and will be done by July 30th. Henceforth, we will try to do better with calibrations.

Item 15: Wipe testing. We have reviewed these procedures and are setting up a revised system whereby all packages will be wipe tested and a record filed with and cross checked against the purchase order. I will investigate any wipe test for which data is not filed by 10 days after receipt of the order.

I hope you will find the above satisfactory.

Yours truly,

Fabian J. Lionetti, Ph.D.
Senior Investigator
Radiation Safety Officer

FJL:ge



8208230283 820818
NMS LIC30
20-04490-01 PDR