



SEWICKLEY VALLEY HOSPITAL, SEWICKLEY, PENNSYLVANIA 15143

July 19,1982

Mr. Thomas T. Martin, Director
U.S. Nuclear Regulatory Commission
Division of Engineering & Technical Programs
Region 1
631 Park Avenue
King of Prussia, Pa. 19406

Dear Mr. Martin:

We have reviewed the violations cited in your letter of June 22,1982. These were the results of an inspection conducted by Dr. Friedman, on May 28,1982.

In order to be in full compliance, we have taken the following action:

<u>VIOLATION</u>	<u>SUBJECT</u>	<u>FINDINGS AND ACTIONS</u>
A	10CFR to appendix C Condition No. 12	A Physician, not named in our license routinely used licensed material. Not applicable because Dr. Goldstein resigned his position here at Sewickley Valley Hospital as of June 25,1982. In the future, we will expedite licensing new radiologists to avoid this problem.
B	10CFR 19.12-in-service including janitorial staff.	We have discussed this violation with Housekeeping Department Head and have arranged a meeting with those individuals whose job area is in the restricted area. Documentation of this meeting will be recorded in Nuclear Medicine Department and in Housekeeping Department for those individuals involved. A special orientation program is being developed for all new housekeeping supervisors and personnel assigned to work in the restricted area.
C-1	Condition. No.17-Use of disposal gloves.	On July 13,1981, Nuclear Medicine Department ordered correct sized and properly fitted gloves from Purchasing Department which will be utilized on all Radialpharmaceutical injections.


8208260123 820823
NMS LIC30
37-11079-01 PDR

C-2 Condition No.17
 Item No.14 Package
 surverys.

Beginning July 6,1982,surverys of packages received into the Nuclear Medicine Department are documented in appropriate log book. This log will include the shipping Number, the date, the activity surveyed, and background information.

We appreciate Dr. L. F. Friedman's review of our procedures and assure you that we have every intention of meeting the Commission's standards.

Very truly yours,



Marvin M. Wedeen,
Assistant Administrator

MMW/jp