Docket No. 30-3790

AUG 2 3 1982

License No. 06-05869-01

Pfizer Incorporated
ATTN: Dr. E. R. Pinsen
Vice President
Central Research
Eastern Point Road
Groton, Connecticut 06340

Gentlemen:

Subject: Inspection No. 82-01

This refers to the routine safety inspection conducted by Mr. C. Stearns of this office on June 30, 1982 of activities authorized by NRC License No. 06-05869-01 and to the discussions of our findings held by Mr. Stearns with yourself and Dr. Figdor of your staff at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, measurements made by the inspector, and observations by the inspector.

Within the scope of this inspection, no violations were observed.

From the discussion of our findings at the conclusion of the inspection, we understand that you plan to occasionally include breathing - zone air sampling during iodinations involving one or more millicuries of iodine-125. We will review this matter during the next inspection of your licensed program.

In accordance with Section 2.790 of the NRC's "Rules of Practice," Part 2, Title 10, Code of Federal Regulations, a copy of this letter will be placed in the Public Document Room.

No reply to this letter is required. Your cooperation with us in this matter is appreciated.

Sincerely,

Original Signed By: John D. Kinneman

Thomas T. Martin, Director
Division of Engineering and Technical
Programs

OFFICIAL RECORD COPY

AUG 2 3 1982

cc:

Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
State of Connecticut

bcc:

Region I Docket Room (with concurrences)
Chief, Operational Support Section (w/o encl)

RI:DETP Stearns/gwc 8/18/82 RI:DETP Kinneman

OFFICIAL RECORD COPY