

APPENDIX A
NOTICE OF VIOLATION

A. Lopez-Rivera, M.D.

License No. 52-16661-01

As a result of the inspection conducted on April 29, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified.

- A. 10 CFR 30.3 requires that no person shall manufacture, produce, transfer, receive, acquire, own, possess, or use byproduct material except as authorized in a specific or general license issued pursuant to NRC Rule and Regulations.

Contrary to the above, the possession and use of the strontium-90 beta eye applicator has been conducted since September 30, 1980 without a valid license.

This is a Severity Level IV Violation (Supplement VI).

- B. 10 CFR 35.14(b)(5) requires that sealed sources containing greater than 100 microcuries of byproduct material be tested for leakage and/or contamination at intervals not to exceed six months.

Contrary to the above, no leak test had been performed on the strontium-90 sealed source within the six month interval prior to the inspection.

This is a Severity Level IV Violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Date: MAY 26 1982