

August 12, 1982

United States Nuclear Regulatory Commission  
Region III  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

Attention: D. J. Sreniawski, Chief  
Materials Radiation Protection  
Section 2

RE: Notice of Violation Response  
License No. 21-17158-02

Gentlemen:

In response to your letter of July 14, 1982, we make the following  
submittal:

VIOLATION #1

1. All gauges have been wipe-tested since May 27, 1982. No problems were found.
2. As a minimum, each gauge will be wipe-tested on the first working day, five months after the most recent wipe-test for that gauge. A calendar will be set-up in the record book for the gauges and the due date for the next scheduled wipe-test marked on the calendar. The calendar will be monitored on a weekly basis to make sure the tests are performed in a timely manner.
3. Full compliance has been achieved as of June 10, 1982.

VIOLATION #2

1. License renewed May 21, 1982 (License #21-17158-02).
2. In December of 1986, we will file for renewal of our existing license. This should provide ample time to complete and process any necessary paperwork.
3. Full compliance was achieved May 21, 1982.

If you have any further questions, please do not hesitate to contact me.

Very truly yours,

SOIL AND MATERIALS ENGINEERS, INC.

*Frank A. Henderson*

Frank A. Henderson  
Radiation Safety Officer

8208300077 820823  
NMS LIC30  
21-17158-02 PDR



soil and materials  
engineers, inc

34400 GLENDALE AVENUE  
LIVONIA, MICHIGAN 48150  
(313) 525-0310

KENNETH W. KRAMER, P.E.  
ROBERT M. HARRELD, P.E.  
JOSEPH D. JOACHIM, P.E.  
RICHARD O. ANDERSON, P.E.  
THEODORE D. DZIURMAN, P.E.  
FRANK A. HENDERSON  
EDWARD S. LINDOW, JR., P.E.