

Appendix

NOTICE OF VIOLATION

Kalamazoo College

License No. 21-13658-01

As a result of the inspection conducted on June 15 and 18, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified:

1. License Condition No. 17 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

The referenced application dated April 9, 1975 states in Item 11 that all gamma-ray sensitive instruments will be calibrated once every year.

Contrary to the above, the Tracerlab SU-14 and Jordon CDV-710 survey meters have not been calibrated since September 1978.

This is a Severity Level IV violation (Supplement VI).

2. License Condition No. 17 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

The referenced letter dated July 30, 1976, states that monthly surveys are to be made of radioactive materials in the storage vault.

Contrary to the above, no surveys have been conducted since September 1978.

This is a Severity Level IV violation (Supplement VI).

3. License Condition No. 14A(1) states each sealed source containing licensed material shall be tested for leakage and/or contamination at intervals not to exceed six months.

Contrary to the above, the Pu-Be neutron source has not been leak tested since September 1978.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full com-

pliance will be achieved. Consideration may be given to extending your response time for good cause shown.

7/2/02  
Dated

D. J. Sreniawski  
D. J. Sreniawski, Chief  
Materials Radiation Protection  
Section 2