

# Edmond Memorial Hospital

One South Bryant  
Edmond, Oklahoma 73034  
Tel. (405) 341-6100

August 9, 1982

RECEIVED  
NRC Personnel  
Region IV

AUG 11 1982

Mr. Glen D. Brown, Chief  
Technical Program Branch  
United States  
Nuclear Regulatory Commission  
Region IV  
611 Ryan Plaza Drive, Suite 1000  
Arlington, Texas 76012

Dear Mr. Brown:

As a result of the routine safety inspection conducted by your office on June 14, 1982, and the subsequent notice of violation, Edmond Memorial Hospital has taken the following corrective steps in the areas where violations were identified.

1. (a) Violation - The survey meter was not calibrated annually as required.

Corrective Action - The survey meter will be calibrated at least annually in accordance with NRC regulations. A back-up survey meter has been ordered.

- (b) Violation - Dose preparation and injection areas were not surveyed daily with a survey meter as required.

Corrective Action - Effective June 15, 1982, daily surveys of the dose preparation and injection areas were begun. Daily surveys will continue with reviews conducted quarterly by the staff radiologist. ✓

- (c) Violation - Records of contamination surveys on incoming packages have not been adequately maintained.

Mr. Glen D. Brown, Chief  
August 9, 1982  
Page 2

Corrective Action - Effective June 15, 1982,  
records of contamination surveys on  
incoming packages are being maintained  
in a log book. The staff radiologist  
will conduct a quarterly review of  
these records.

2. Violation - Sealed calibration sources have not been tested for leakage and/or contamination at least every six months.


Corrective Action - The following sealed calibration sources were tested for leakage and/or contamination on June 15, 1982:

Co-57 2.77 mCi  
Ba-133 231 uCi  
Cs-137 190 uCi

Sealed calibration sources will be tested at least every six months. A log will be maintained on tests conducted. The staff radiologist will review the log quarterly.

Should you have any question regarding corrective actions taken, please feel free to call.

Sincerely,



Joel A. Hart  
Administrator