

Appendix

NOTICE OF VIOLATION

GTE Automatic Electric, Inc.

License No. 12-15845-01

As a result of the inspection conducted on June 29, 1982, and in accordance with the NRC Enforcement Policy, 47 CFR 9987 (March 9, 1982), the following violation was identified:

License Condition No. 20 requires that your chromatograph detector containing nickel-63 be tested for leakage and/or contamination at intervals not to exceed six months.

Contrary to this requirement, you have failed to test your chromatograph detector containing nickel-63 for contamination or leakage since September 1981, a period greater than six months.

This is a Severity Level V violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated 7/7/82

D. G. Wiedeman, Chief
Materials Radiation Protection
Section 1