

26 AUG 1982

Docket No. 30-11991

License No. 20-16970-01

Martha's Vineyard Hospital
ATTN: Mr. David Federowicz
Executive Director
Box 1477
Oak Bluffs, Massachusetts 02557

Gentlemen:

Subject: Inspection No. 30-11991/82-01

This refers to your letter dated August 3, 1982, in response to our letter dated July 7, 1982.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

From an August 12, 1982, telephone conversation between yourself and Dr. L. Friedman of this office, we understand that your dose calibrator will be checked daily with your cesium-137 source on all buttons (isotope settings) to be used that day, that dose calibrator linearity will be checked quarterly, and that your Chief Radiological Technician will perform monthly audits of your licensed program and report the results at the monthly meeting of your Isotopes Committee. Please inform this office if your intent differs from our understandings.

Your cooperation with is appreciated.

Sincerely,
Original Signed By:

Jack Davis

W Thomas T. Martin, Director
Division of Engineering and
Technical Programs

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
Commonwealth of Massachusetts
Olivia Reeney, R. T., Chief Radiological Technician
Peter O'Hara, M. D., Radiologist

bcc:
Region I Docket Room (with concurrences)
RI:DETP RI:DETP
Friedman/dmg Kinneman
8/19/82

JF *8/24/82*
8208310259 820826
NMS LIC30
20-16970-01 PDR

OFFICIAL RECORD COPY

1E07