



St Joseph's Hospital and Medical Center

703 Main Street - Paterson - New Jersey 07503 - (201) 977-2040

182 AED 27 M1:23

Office of the Executive Vice President

CFFICE OF SECRETARY DOCKETING & SERVICE BRANCH

August 23, 1982

OPERS HUMBER PR-Misc. Notice

Reg Guide)

Secretary of the Commission U. S. Nuclear Regulatory Commission Washington, D. C. 20555

Attention: Docketing and Service Branch

Dear Sir:

We have reviewed the Draft Regulatory Guide TM 608-4 you sent us on July 7, 1982. We find that we can accept as written all regulations with the exception of Item 13 C.

Item 13 C specifies that dosimetry systems be calibrated every two years. We suggest that this requirement be relaxed to accommodate the long time required to get instruments calibrated by the rather few accredited calibration laboratories. Since it takes a year or more to obtain calibration service, the two year requirement seems to be impractical.

Sincerely yours,

Executive Vice President

PCW/vmp

8208310267 820823 PDR REGGD 820823

add: Ed Hill 5650 NL

Acknowledged by card. 8/30/82 engl