

AUG 24 1982

Fairview General Hospital
ATTN: Mr. Kenneth Misener
Associate Administrator
18101 Lorain Avenue
Cleveland, Ohio 44111

30-2699
License No. 34-02162-01

Gentlemen:

This refers to the routine safety inspection conducted by Mr. W. J. Slawinski of this office on July 28, 1982, of activities authorized by NRC Byproduct Material License No. 34-02162 and to the discussion of our findings with you and other members of your staff at the conclusion of the inspection.

This also acknowledges receipt of your letters dated August 5, 1982, concerning the high film badge reading during the period January 20, 1980, through February 19, 1980.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and with the conditions of your license. The inspection consisted of a selective examination of procedures and representative records, observations, independent measurements, and interviews with personnel.

During this inspection, certain of your activities appeared to be in non-compliance with NRC requirements, as specified in the enclosed Appendix. A written response is required.

Regarding your letters of August 5, 1982, it appears the whole body film badge and wrist badge readings of 4.27 rem and 3.38 rem respectively for the period January 20, 1980 through February 19, 1980, occurred from x-ray and not from improper use of licensed material. Although we have not identified these badge exposures as items of noncompliance we are concerned that these high film badge readings went unnoticed until brought to your attention by our inspector. Therefore, describe what actions you will take to assure us that such areas are not overlooked in the future.

IE HQ FILE COPY

8208310407 820824
NMS LIC30
34-02162-01 PDR

DMB IE-07

Fairview General Hospital

2

AUG 24 1982

The responses directed by this letter (and the accompanying Notice) are not subject to the clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, PL 96-511.

We will gladly discuss any questions you have concerning this inspection.

Sincerely,

Original Signed by J. R. Miller

J. R. Miller, Chief
Technical Inspection Branch

Enclosure: Appendix,
Notice of Violation

cc w/encl:
DMB/Document Control Desk (RIDS)

RFIN
Slawinski/sv
8/19/82

RIII
Sreniawski

RILL
Miller
8/24/82