

JUL 21 1982

Docket No. 30-11242

License No. 20-01476-05

The Gillette Company
ATTN: Mr. J. Galligan
Vice-President, Research
and Development
Personal Care Division
T Building
Gillette Park
Boston, Massachusetts 02106

Gentlemen:

Subject: Inspection No. 30-11242/82-01

This refers to your letter dated June 28, 1982, in response to our letter dated June 11, 1982.


Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

Please be advised that your Ni-63 source requires leak testing once every six (6) months rather than on a monthly basis as you indicated in your corrective actions.

Your cooperation with us is appreciated.


Sincerely,


Original Signed By:
John D. Kinneman

 Thomas T. Martin, Director
Division of Engineering and
Technical Programs

cc:
Mr. R. Sneath, Manager of Chemical
Research and Radiation Safety Officer
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
Commonwealth of Massachusetts (2)

bcc:
Region I Docket Room (with concurrences)

 RI:DETP
Johansen/lp
7/12/82

 RI:DETP
Kinneman

7/20/82

OFFICIAL RECORD COPY

IE:07

8207290214 820721
NMS LIC30
20-01476-05 PDR