

APPENDIX A

NOTICE OF VIOLATION

St. Francis Central Hospital
Pittsburgh, Pennsylvania 15219

Docket No. 030-10656
License No. 37-16245-01

During an NRC inspection conducted on December 15, 1993, a violation of NRC requirements was identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, the violation is listed below:

10 CFR 35.50(c) requires, in part, a licensee to perform appropriate tests for linearity and geometry dependence required by 10 CFR 35.50(b) following adjustments or repair of the dose calibrator.

Contrary to the above, following the adjustment of the dose calibrator on May 20, 1991, the licensee did not perform a test for linearity or geometry dependence.

This is a Severity Level IV violation. (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, St. Francis Central Hospital is hereby required to submit a written statement or explanation to the U.S. Nuclear Regulatory Commission, ATTN: Document Control Desk, Washington, D.C. 20555, with a copy to the Regional Administrator, Region I, within 30 days of the date of the letter transmitting this Notice of Violation (Notice). This reply should be clearly marked as a "Reply to a Notice of Violation" and should include for each violation: (1) the reason for the violation, or, if contested, the basis for disputing the violation, (2) the corrective steps that have been taken and the results achieved, (3) the corrective steps that will be taken to avoid further violations, and (4) the date when full compliance will be achieved. If an adequate reply is not received within the time specified in this Notice, an order or a Demand for Information may be issued to show cause why the license should not be modified, suspended, or revoked, or why such other action as may be proper should not be taken. Where good cause is shown, consideration will be given to extending the response time.

OFFICIAL RECORD COPY - S:\PENDING\STFRAN.NOV - 01/12/94

9402170082 940112
PDR ADOCK 03010656
C PDR

RETURN ORIGINAL TO
REGION I

IE:07