



Kenosha Memorial Hospital

6308 EIGHTH AVENUE • KENOSHA, WI 53140-5082

JOHN C. McGINTY, JR., President
Telephone • 414-656-2011

SINCE 1903

August 19, 1982

Mr. D. G. Wiedeman, Chief
Materials Radiation Protection, Sec. 1
U. S. Nuclear Regulatory Commission, Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Dear Mr. Wiedeman:

We appreciate the thorough inspection covered by Mr. Burgin, radiation specialist, on July 21, 1982.


The purpose of this letter is to advise you of the steps which have been taken to assure present and future compliance with the recommendations from Mr. Burgin. They are:

- Monthly linearity checks of our dose calibrator were initiated as of August 4, 1982.
CS 137 source obtained and weekly constancy checks initiated as of August 3, 1982.
- Daily GM surveys being done in hot lab as of July 22, 1982. Weekly wipe tests of removable contamination being done weekly as of July 23, 1982.
- A calendar has been established by Dr. Connolly, the radiation safety officer, to assure that the committee meets at least quarterly.
- Disposal records of decayed waste now being documented as of August 2, 1982.

We trust this brings us in complete compliance. Thank you again for your efforts in our behalf.

Very truly yours,

KENOSHA MEMORIAL HOSPITAL


Peter M. Friend
Vice President
Administration

PMF/em
cc: Dr. Yerex/Lynn Poker

8209030239 820827
NMS LIC30
48-00056-02 PDR

AUG 24 1982