

SEP 0 2 1982

30-2671

Lakewood Hospital
ATTN: Mr. Harry J. Vandavelde
Assistant Administrator
14519 Detroit Avenue
Lakewood, OH 44107

License No. 34-01197-01

Gentlemen:

Thank you for your letter dated August 18, 1982, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated July 21, 1982.


This letter also refers to the telephone discussion on August 31, 1982, between Mr. D. G. Wiedeman, of my staff, and Mr. P. Early, of Nuclear Medicine Associates, concerning the assaying of xenon doses in a dose calibrator. It is our understanding that you will send in an amendment request to add to your license the procedures for determining xenon doses as described in your letter dated August 18, 1982.

If your understanding differs from the above, please contact Darrel Wiedeman at 312-932-2500.

We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,


J. R. Miller, Chief
Technical Inspection Branch

cc w/ltr dated 8/18/82:
William J. Fayen, M.D.
Mr. Ron Amrich, Department
Head Biometrix
DMB/Document Control Desk (RIDS)

8209080609 820902
NMS LIC30
34-01197-01 PDR

IE-HO. File Copy
DMB IE-07

OFFICE	RIII	RIII	RIII	RIII		
SURNAME	Reichhold/cw	Wiedeman	Mallett	Miller		
DATE	9/1/82	9/1/82	9/1/82			