

AUG 24 1982

APPENDIX A

NOTICE OF VIOLATION

Formosa Plastics Corporation, Delaware
Delaware City, Delaware 19706

Docket No. 030-19393
License No. 07-20540-01

As a result of the inspection conducted on July 15, 1982, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violation was identified:

Condition 14 of License No. 07-20540-01 requires that installation, relocation, maintenance, repair, and initial radiation survey of devices containing byproduct material be performed only by persons specifically authorized by the Commission or an Agreement State to perform such services.

Contrary to the above, two gauges, each containing 20 millicuries of cesium-137, were installed on tanks PT 301A and PT 301B by your employees, who do not have the required authorization from the Commission or an Agreement State to install such devices.

This is a Severity Level IV violation. (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, Formosa Plastics Corporation, Delaware, is hereby required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including the corrective steps which have been taken and the results achieved, corrective steps which will be taken to avoid further violations, and the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

OFFICIAL RECORD COPY

8209090126 820824
NMS LIC30
07-20540-01 PDR