



**soil and materials
engineers, inc**

34400 GLENDALE AVENUE
LIVONIA, MICHIGAN 48150
(313) 525-0310

KENNETH W. KRAMER, P.E.
ROBERT M. HARRELD, P.E.
JOSEPH D. JOACHIM, P.E.
RICHARD O. ANDERSON, P.E.
THEODORE D. DZURMAN, P.E.
FRANK A. HENDERSON
EDWARD S. LINDOW, JR., P.E.

August 12, 1982

United States Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Attention: D. J. Sreniawski, Chief
Materials Radiation Protection
Section 2

RE: Notice of Violation Response
License No. 21-17158-02

Gentlemen:

In response to your letter of July 14, 1982, we make the following
submittal:

VIOLATION #1

1. All gauges have been wipe-tested since May 27, 1982. No problems were found.
2. As a minimum, each gauge will be wipe-tested on the first working day, five months after the most recent wipe-test for that gauge. A calendar will be set-up in the record book for the gauges and the due date for the next scheduled wipe-test marked on the calendar. The calendar will be monitored on a weekly basis to make sure the tests are performed in a timely manner.
3. Full compliance has been achieved as of June 10, 1982.

VIOLATION #2

1. License renewed May 21, 1982 (License #21-17158-02).
2. In December of 1986, we will file for renewal of our existing license. This should provide ample time to complete and process any necessary paperwork.
3. Full compliance was achieved May 21, 1982.

If you have any further questions, please do not hesitate to contact me.

Very truly yours,

SOIL AND MATERIALS ENGINEERS, INC.

Frank A. Henderson

Frank A. Henderson

Radiation Safety Officer

TECHNICAL AND MATERIALS CONSULTANTS

*IF HQ file copy
DMB IE-07*

AUG 16 1982

8209150219 820812 PDR
NMS LIC30
21-17158-01

bh/