

SEP 07 1982

30-394

The Cleveland Clinic Foundation  
ATTN: James Krieger, M.D.  
Executive Assistant of  
the Board of Governors  
9500 Euclid Avenue  
Cleveland, Ohio 44106

License No. 34-00466-02

Gentlemen:

Thank you for your letter dated August 12, 1982, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated July 21, 1982.

Regarding item 2 of your response letter dated July 21, 1982, our records show that the full calibration of your teletherapy unit performed on January 30, 1982, was made using a dosimetry system that was not calibrated as required by 10 CFR 35.23(a). Therefore, we conclude that this matter constituted an item of noncompliance. We will examine your corrective actions during a future inspection.

Your cooperation with us is appreciated.

Sincerely,

Charles E. Norelius, Director  
Division of Engineering and  
Technical Programs

cc w/ltr dtd 8/12/82:  
DMB/Document Control Desk (RIDS)

8209150259 820907  
NMS LIC30  
34-00466-02 PDR

IE HQ  
DMB IE-07

|         |             |            |        |          |  |  |
|---------|-------------|------------|--------|----------|--|--|
| OFFICE  | RIII        | RIII       | RIII   | RIII     |  |  |
| SURNAME | Mullauer cw | Sreniawski | Miller | Norelius |  |  |
| DATE    | 9/2/82      |            |        | 9/3/82   |  |  |