

JUL 15 1982

APPENDIX A
NOTICE OF VIOLATION

Jefferson Memorial Hospital

License No. 45-17911-01

As a result of the inspection conducted on July 2, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified.

- A. 10 CFR 20.105(b)(2) requires that no licensee shall possess or use licensed material in such a manner as to create in any unrestricted area, radiation levels which, if an individual were continuously present in the area, could result in his receiving a dose in excess of 100 millirems in any seven consecutive days.

Contrary to the above, radiation levels in the unrestricted patient waiting area outside the Nuclear Medicine Department were 1.0 millirems per hour, so that an individual, if continuously present, could receive a dose in excess of 100 millirems in any seven consecutive days.

This is a Severity Level IV Violation (Supplement IV).

- B. License Condition 15 requires that licensed material be used in accordance with statements and procedures contained in your application dated December 15, 1977. Item 10 of Enclosure 6A to your application prohibits eating or drinking in the radioisotope work area.

Contrary to the above, during the inspection of July 2, 1982, several unsealed containers of consumable beverages were observed under the sink in the radioisotope work area.

This is a Severity Level V Violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Date: _____