

APPENDIX A
NOTICE OF VIOLATION

Jose T. Medina, M.D.

License No. 52-14931-01

As a result of the inspection conducted on September 2, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified.

- A. License Condition 17 requires that material be used in accordance with statements and procedures contained in your application dated November 30, 1976, and letter dated March 7, 1977. One procedure in your application states that elution and dose preparation areas shall be surveyed for contamination daily.

Contrary to the above, your elution and dose preparation areas are surveyed for contamination at weekly intervals rather than daily.

This is a Severity Level V Violation (Supplement VI).

- B. License Condition 17 requires that material be used in accordance with statements and procedures contained in your application dated November 30, 1976, and letter dated March 7, 1977. One statement in your letter requires that daily constancy checks be performed on your dose calibrator.

Contrary to the above, the constancy of your dose calibrator is being checked on a weekly basis rather than daily.

This is a Severity Level V Violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Date: September 13, 1982

8210150050 821007
NMS LIC30
52-14931-01 PDR