

NOTICE OF VIOLATION

V. A. Harry S. Truman
Memorial Veterans Hospital
Columbia, Missouri

Docket No. 030-13378
License No. 24-15235-03

As a result of the inspection conducted on July 11, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1990) (Enforcement Policy) the following violation was identified:

License Condition 23.A. requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced documents.


The referenced application dated February 10, 1989, states in Item 8 that training for supervisory nursing personnel will be given at least once each year.

Contrary to the above, supervisory nursing personnel had not been given training at least once each year as required. Specifically, training to supervisory nursing personnel had not been given since September 1988.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

August 10, 1990
Dated


R. J. Caniano, Chief
Nuclear Materials Safety
Section 2