

NOTICE OF VIOLATION

Central Wisconsin Engineers
Rothschild, WI

License No. 48-24978-01
Docket No. 030-30040

As a result of the inspection conducted on August 28, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violations were identified:

1. License Condition 17 requires that all licensed material be possessed and used in accordance with statements, representations and procedures contained in application dated May 23, 1987.

Application dated May 23, 1987, states that when a moisture/density gauge is transported in an open bed vehicle, the gauge should be fastened and locked to the bed of the vehicle.

Contrary to the above, on August 28, 1990, a Troxler moisture/density gauge was transported in an open bed truck, but it was not locked to the bed of the truck.

This is a Severity Level IV violation (Supplement VI).

2. License Condition 12.A(1) states that sealed sources shall be tested for leakage and/or contamination at intervals not to exceed 6 months.

Contrary to the above, the cesium-137 and americium-241 sealed sources in the Troxler moisture/density gauge (serial 14785) had not been tested for leakage and/or contamination in July 1990.

This is a Severity Level IV violation (Supplement VI).

3. License Condition 14 states that a physical inventory to account for all gauges will be performed every 6 months.

Contrary to the above, a physical inventory to account for the Troxler moisture/density gauge (serial 14785) was not performed in July 1990.

This is Severity Level IV violation (Supplemental VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

September 13, 1990
Dated

George M. McCann
George M. McCann, Chief
Nuclear Materials Safety
Section 1

9009280172 900913
REG3 LIC30
48-24978-01 PDC