

SACRED HEART HOSPITAL
AND
REHABILITATION CENTER



1430 DEKALB ST., NORRISTOWN, PA.
PHONE: 278-8200

1 May 1990

REGION 1
50
MOHAMED M. SHANBAKY, CHIEF
Nuclear Materials Safety Section A
Division of Radiation Safety and Safeguards
U.S. NUCLEAR REGULATORY COMMISSION, Region 1
475 Allendale Rd.
King of Prussia, PA 19406

Re: License No. 37-01563-03
Subject: Inspection No. 030-02989/90/001

Dear Mr. Shanbaky:

This is in response to your letter dated 21 Apr 90, regarding the above inspection.

In order to assure compliance with Appendix D, Item C and Appendix I, Item 1 in Regulatory Guide 10.8, Revision 1b, we have initiated records for dose calibrator constancy testing for weekends and holidays when the dose calibrator is used.

We have also instituted the use of daily room survey sheets to be completed after patient imaging which may take place on weekends and holidays.

Full compliance was achieved on 23 Mar 90 and is ongoing.

Yours truly,

ARTHUR GODIN, President
NICHOLAS R. RORICK, M.D., RSO

ACG/nrr

9010040204 900918
REG1 LIC30
37-01563-03 PNU



A NOT FOR PROFIT COMMUNITY HOSPITAL OWNED AND OPERATED BY
THE MISSIONARY SISTERS OF THE MOST SACRED HEART OF JESUS