



Jackson County
Schneck Memorial Hospital

200 SOUTH WALNUT STREET
SEYMOUR, INDIANA 47274

December 17, 1982

Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, IL 60137

ATTENTION: Mr. Sreniawski

Dear Mr. Sreniawski:

This letter is in response to your letter dated 11-16-82.


Pursuant to provision of 10 CFR 2.201 quarterly physical inventories for all sources will be performed by the Radiation Safety Officer.

The Radiation Safety Officer will report the results directly to the Radiation Safety Committee. This will insure immediate compliance with the NRC regulations.


Enclosed is a copy of the physical inventory which was performed on 11-17-82.

Sincerely,


Mr. George James, Hospital Administrator


Peter Kamperschroer, M.D.

Richard Stewart, Radiation Safety Officer


Richard Stewart RT RSO

jb

Enclosure

JACKSON COUNTY SCHNECK MEMORIAL HOSPITAL

10 CFR Part 35
RADIONUCLIDE (SEALED SOURCE) INVENTORY

Radionuclide =====	Serial No. =====	Calibration Data =====	Activity on 11-17-82	Location on 11-17-82
Dose Calibrator Sources: -----				
Cesium-137	229-292-5	201 uCi on 6/4/76	168.1 uCi	cabinet in imaging room
Barium-133	229-88-3	234 uCi on 2/17/76	155 uCi	"
Cobalt-57	2060781A-04	5.2 mCi on 7/23/81	1.52 mCi	"
Cobalt-60	229-83-12	54 uCi on 2/12/76	23.5 uCi	"
Cobalt-57	231-010-12	355 uCi on 6/8/76	.9 uCi	hot lab
Cobalt-57	2060579B-14	5 mCi on 5/29/79	199.2 uCi	hot lab
Flood Sources: =====				
Cobalt-57	1295	3.0 mCi on 10/76	10.5 uCi	hot lab
Cobalt-57	615	2.0 mCi on 10/74	1.1 uCi	hot lab
Cobalt-57	2930382C-02	2 mCi on 3/15/82	1.07 uCi	imaging room
Mo-99 Equivalent Sources: =====				
Cesium-137	29.9 Mo eq	4.3 uCi on 10/21/81	4.1 uCi	hot lab
Cesium-137	80230281A-07	31.0 uCi on 2/2/81	29.8 uCi	hot lab
Cesium-137	109.3 Mo eq	18.2 uCi on 10/21/81	16.7 uCi	hot lab
Cesium-137	697-183-30	22 uCi on 3/2/78	19.4 uCi	hot lab
Cesium-137	102.2 Mo eq	16.4 uCi on 10/21/81	16.0 uCi	hot lab
Miscellaneous Sources: =====				
Cobalt-57	button	9.5 uCi on 7/15/74		hot lab
Cobalt-60	dosimeter cal.	10 uCi		hot lab

All sources accounted for:

R. S. Stewart 11-17-82
RSC