



SAGINAW GENERAL HOSPITAL

1447 N. Harrison
Saginaw, Michigan 48602
Tel. (517) 771-4000

December 20, 1982

P. J. Sreniawski, Chief
Materials Radiation Protection, Section 2
U.S. Nuclear Regulatory Commission, Region III
Glen Ellyn, IL 68137

RE: License #21-00607-04

Dear Sir:

In response to your letter of December 7, 1982, following inspection of our facilities on November 19, 1982, the following is submitted for your approval:

#1 & 2. Instructions have been issued to the Radiation Safety Officer to make or cause to make surveys in unrestricted areas including hallways adjacent to the rooms of patients undergoing I-131 therapy. The area survey forms used for recording these proceedings have been modified accordingly.

#3. Full compliance with 10 CFR 20.201 (b) and 105 b will be achieved at the time of the performance of our next radioactive therapy procedure.

Sincerely,

A handwritten signature in cursive script that reads "W. J. Heath". The signature is written in dark ink and is positioned above the printed name and title.

William J. Heath
Vice President

nkr

cc: Ronald Siwik, M.D.
Director of Radiology

DEC 27 1982

8301040367 821230
NMS LIC30
21-00607-04 PDR