



CONSUMERS
POWER
COMPANY

D P Hoffman
Plant Superintendent

Big Rock Point Nuclear Plant, Box 591, Route 3, Charlevoix, MI 49720

October 29, 1982

Nuclear Regulatory Commission
Al Januska
799 Roosevelt Road
Glen Ellyn, Ill 60137

Dear Mr. Januska;

The spiked air sample for confirmatory analysis sent from your office on October 20, 1982 was analyzed on our NP6685 System. It was counted on the one centimeter shelf with the side having the X down. Results were decayed back to March 8, 1979.

Our results gave the following μ ci content:

Co - 60	2.34E-03 μ ci
Cs - 134	1.49E-03 μ ci
Cs - 137	4.27E-03 μ ci
Co - 57	6.58E-04 μ ci

This should complete the addendum to I&E Inspection Report 50-155/82-07 and close item 50-155/82-07-02.

Sincerely,

Gill Fox
Chemistry & Radiation Protection