

MAR 03 1994

SCDB:SCO,
71-0354

Advanced Medical Systems, Inc.
ATTN: Ms. Sherry J. Stein
Director, Regulatory Affairs
121 North Eagle Street
Geneva, Ohio 44041

Dear Ms. Stein:

SUBJECT: NRC INSPECTION REPORT NO. 710354/93-06
NOTICE OF VIOLATION, NOTICE OF NONCONFORMANCE

This refers to your response dated September 28, 1993, concerning the Notice of Violation and Notice of Nonconformance identified in U.S. Nuclear Regulatory Commission Inspection Report No. 710354/93-06; and your request for amendment to your Quality Assurance Program Approval No. 0354, dated February 17, 1994.

The corrective actions which have been or will be taken regarding the Notice of Violation and Notice of Nonconformance are acceptable, and no additional information is requested. However, please note that we expect you to evaluate the replacement parts lists for your transportation packagings to establish the importance to safety of each component, and to control the procurement of those parts accordingly.

The corrective actions taken and the implementation of your Quality Assurance Program will be reviewed in future inspections of your transportation activities.

Sincerely,

Original Signed by

Robert L. Baer, Chief
Source Containment and Devices Branch
Division of Industrial and
Medical Nuclear Safety, NMSS

08-152

Distribution:

NRC F/C ✓
SCDB r/f
STreby, OGC
DReid

IMNS c/f
CPaperiello
RBangart, OSP
SBaggett

NRC PDR
WBrach
DWeiss, OC
CChaughey

NMSS r/f
JLieberman, OE
CNorelius, RIII

OFC	SCDB	C	JJK SCDB	C	SCDB				
NAME	SO Connor		JJankovich		RBaer				
DATE	03/02/93		3/2/93		3/13/93				

C = COVER

E = COVER & ENCLOSURE

N = NO COPY

NT03/10

9403150286 940303
PDR ADDCK 07100354
C PDR

NRC FILE CENTER COPY