

JAN 04 1983

Radiology Associates, P.C.
ATTN: John D. Mellen, M.D.
1412 East 11 Mile Road
Royal Oak, MI 48067

License No. 21-16754-01

Gentlemen:

Thank you for your letter dated December 22, 1982, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated December 20, 1982. We will examine these matters during a future inspection.

In reference to the location of the dose calibrator put in use without a test for geometrical variation, we understand that the unit was at 11885 E. 12 Mile Road not 16800 W. 12 Mile Road as we had originally stated and that the test has been made and all appears now in order.

Your cooperation with us is appreciated.

Sincerely,

D. J. Sreniawski, Chief
Materials Radiation Protection
Section 2

cc w/ltr dtd 12/22/82:
DMB/Document Control Desk (RIDS)

8301110145 830104
NMS LIC30
21-16754-01 PDR

IE H.O. File Copy
DMB IE-07

OFFICE	RIII <i>DM</i>	RIII <i>DM</i>	RIII <i>DM</i>				
SURNAME	Douglas/cw	Matson <i>1/4</i>	Sreniawski				
DATE	1/4/83		1/4/83				