

DMB

January 6, 1983

Docket No. 50-155

Consumers Power Company  
ATTN: R. B. DeWitt  
Vice President  
Nuclear Operations  
212 West Michigan Avenue  
Jackson, MI 49201

Gentlemen:

Attached is an addendum, Table II, to Report No. 50-155/82-07 transmitted by our May 20, 1982, letter. All comparative beta analyses results of the liquid sample collected during the April 27-29, 1982, inspection are in agreement. These comparisons close Open Item 50-155/82-07-01.

A spiked air particulate filter was analyzed following recalibration of the site's analytical system. All four comparisons were categorized as agreements or possible agreements; this closes Open Item 50-155/82-07-02.

Please attach this addendum to your copy of the report. We will gladly discuss any questions you have concerning this inspection.

Sincerely,

C. J. Paperiello, Chief  
Emergency Preparedness and  
Program Support Branch

Attachments: Table II

cc w/encl:  
D. J. Vandewalle, Nuclear  
Licensing Administrator  
D. P. Hoffman, Plant  
Superintendent  
DMB/Document Control Desk (RIDS)  
Resident Inspector, RIII  
Ronald Callen, Michigan  
Public Service Commission

RIII  
*RM*  
Nicholson/jp  
1/4/83

RIII  
*SR*  
Rozak

RIII  
*SR*  
Schumacher

RIII  
*CJP*  
Paperiello  
1/4/83

RIII  
*AB*  
Boyd

8301110452 830106  
PDR ADOCK 05000155  
G PDR

IE:06