Appendix

NOTICE OF VIOLATION

Distribution Enterprises, Incorporated

License No. 12-20176-01

As a result of the inspection conducted on December 28, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified:

 License Condition No. 13 requires that you test your sealed sources containing byproduct material for leakage and/or contamination at intervals not to exceed six months.

Contrary to this requirement, as of the day of the inspection, December 28, 1982, you had not tested your sealed source containing americium-241 for contamination or leakage since September 4, 1981, an interval of more than six months.

This is a Severity Level IV violation (Supplement VI).

2. 10 CFR 19.11(a) and (b) require that current copies of Part 19, Part 20, your license, license conditions, documents incorporated into the license, license amendments and operating procedures be posted, or that a notice describing these documents and where they may be examined, be posted.

Contrary to this requirement, on the day of the inspection, December 28, 1982, neither the documents nor the notices were posted.

This is a Severity Level V violation (Supplement VI).

With respect to Item No. 2, the inspection showed that action had been taken to correct the identified item of noncompliance and to prevent recurrence. Consequently, no reply to this item of noncompliance is required and we have no further questions regarding this matter. With respect to Item No. 1, pursuant to the provisions of 10 CFR 2.201 you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

D. G. Wiedeman, Chief

Materials Radiation Protection

Section 1