

001003

MAR 14 1994

John R. Sinkey, M.D.
3036 Sylvania Avenue
Toledo, OH 43613-4128

License No. 34-02867-01
Docket No. 030-02710

Dear Dr. Sinkey:

SUBJECT: SATISFACTION OF CONFIRMATORY ACTION LETTER NO. CAL-RIII-93-013,
DATED JULY 28, 1993

This refers to your letters dated September 24, 1993, and February 16, 1994,
in response to our Confirmatory Action Letter (CAL) dated July 28, 1993. The
CAL confirmed that you would not resume licensed activities until you had:

1. Obtained the services of a technically qualified person(s) who is sufficiently knowledgeable to properly implement your radiation safety procedures.
2. Repaired or replaced, and fully calibrated, your dose calibrator.
3. Submitted a revised Quality Management Program that meets the specific objectives contained in 10 CFR 35.32.

Based on the information provided, we have determined that you have satisfied the terms of the CAL and we have no further questions at this time.

In accordance with 10 CFR 2.790 of the Commission's regulations, a copy of this letter will be placed in the NRC Public Document Room.

Sincerely,

ORIGINAL SIGNED W. L. AXELSON

9403220012 940314
PDR ADDCK 03002710
C PDR

W. L. Axelson, Director
Division of Radiation Safety
and Safeguards

bcc: H. Thompson, DEDO
J. Lieberman, OE
J. Goldberg, OGC
R. Bernero, NMSS
J. Glenn, NMSS
J. Martin, RIII
R. DeFayette, RIII
PUBLIC - IE07

RIII <i>Casey</i> Casey/dsc 03/10/94	RIII <i>Frazier</i> Frazier 03/10/94	RIII <i>Holt</i> Holt 03/10/94	RIII <i>Caniano</i> Caniano 03/10/94	RIII <i>Berson</i> Berson 03/10/94	RIII <i>DeFayette</i> DeFayette 03/10/94	RIII <i>Axelsson</i> Axelsson 03/14/94
-----------------------------------------------	-----------------------------------------------	-----------------------------------------	-----------------------------------------------	---------------------------------------------	---------------------------------------------------	-------------------------------------------------

210038

110
FE07