

BLOOMINGTON HOSPITAL  
P.O. Box 1149  
Bloomington, Indiana 47402

January 28, 1983

D. J. Sreniawski, Chief  
Materials Radiation Protection  
Section 2  
U.S. Nuclear Regulatory Commission  
Region III  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

Re: Inspection response license #13-10408-02


Dear Mr. Sreniawski:

This letter is in reference to the official NRC inspection report our facility (license #13-10408-02) received on January 6, 1983. The following is the response to the three items of non-compliance cited during the December 20 and 21, 1982 inspections. The responses are in the same order as the official report.

1. An amendment has now been prepared notifying your agency of the move and planned use of Xe-133 for ventilation studies. This amendment is currently pending, and we confirm no Xe-133 ventilation studies will be conducted prior to receiving the approved amendment.
2. Wipe testing of Mo99/Tc99m generators prior to shipping has now been initiated. We confirm wipe test procedures and action levels for decontamination will follow procedures outlined by our original NRC license application, Regulatory Guide 10.8 and 49CFR 173.397.
3. We confirm wipe testing of incoming packages containing radioactive materials has been initiated. Procedures described in Appendix F of the NRC Draft license guide, dated November 1977, will be followed.

We feel the steps outlined above indicate appropriate action has been taken to ensure full compliance with our NRC radioactive materials license. If there are any further questions, please feel free to contact us.

Sincerely,

  
L. Gene Perry  
Vice President - Professional Services

/jmk

8302150039 830208  
NMS LIC30  
13-10408-02 PDR

FEB 3 1983