

FEB 1 1983

APPENDIX A

NOTICE OF VIOLATION

Kaminski Brothers, Inc.
Avoca, Pennsylvania 18641

Docket No. 30-19375
License No. 37-20545-01

As a result of the inspection conducted on January 12, 1983, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violations were identified:

- A. 10 CFR 20.105(b) requires that radiation levels in unrestricted areas be limited so that an individual who was continuously present in the area could not receive a dose in excess of 2 millirems in any hour or 100 millirems in any seven consecutive days.

Contrary to the above, on January 12, 1983, radiation levels of 0.6 millirem per hour (100 millirem in seven days) existed 18 inches from the Troxler gauge stored in the Engineering Laboratory, an unrestricted area, and had existed at that level for a month.

This is a Severity Level IV violation (Supplement IV).

- B. Condition 17 of License No. 37-20545-01 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in an application dated March 18, 1982.

Supplement No. 4 to this application requires that the Surface Moisture Density Gauge be stored in a locked cabinet in the basement of the scale house.

Contrary to the above, as of January 12, 1983, the gauge was not stored in a locked cabinet, nor had the required cabinet been constructed.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, Kaminski Brothers, Inc., is hereby required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including; (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

8302160276 830201
NMS LIC30 PDR
37-20545-01