

SAFETY INSPECTION

1. LICENSEE Massachusetts General Hospital 1 Fruit Street Boston, Massachusetts 02114		2. REGIONAL OFFICE REGION I U S NUCLEAR REGULATORY COMMISSION 475 ALLENDALE ROAD KING OF PRUSSIA PA 19406-1415	
3. DOCKET NUMBER(S) 030-00948	4. LICENSE NUMBER(S) 20-03814-81	5. DATE OF INSPECTION March 1-2, 1994	

LICENSEE:
The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

1. Within the scope of this inspection, no violations were observed.

2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.

3. During this inspection certain of your activities, as described below or attached, were in violation of NRC requirements. This form is a NOTICE OF VIOLATION, which is required to be posted in accordance with 10 CFR 19.11.

A. _____ was not properly posted to indicate the presence of a _____ 10 CFR 20.203(b),(c),(d),(e) or 34.42.

B. _____ of sealed sources were not performed at the proper frequencies. 10 CFR _____ or License Condition Number _____.

C. Records of _____ were not properly maintained. 10 CFR _____ or License Condition Number _____.

D. Documents were not properly posted or otherwise made available. 10 CFR 19.11.

E. Reports or notification of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____.

F. _____

DESIGNATED ORIGINAL
Certified by: M. A. [Signature]

9403310012 940302
PDR ADDCK 03008948
C PDR

FE07
110

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

SIGNATURE - LICENSEE 300043	DATE	SIGNATURE - NRC INSPECTOR <u>Richard [Signature]</u>	DATE 3/17/94
------------------------------------	------	---	-----------------

SAFETY INSPECTION

1. LICENSEE Massachusetts General Hospital 1 Fruit Street Boston, Massachusetts 02114		2. REGIONAL OFFICE REGION I U S NUCLEAR REGULATORY COMMISSION 475 ALLENDALE ROAD KING OF PRUSSIA PA 19406-1415	
3. DOCKET NUMBER(S) 030-01948	4. LICENSE NUMBER(S) 20-0381A-R1	5. DATE OF INSPECTION March 1-2, 1994	

3. (Continued)

G. _____

H. _____

I. _____

4. The violations listed below are not being cited because they were self-identified, and corrective action was or is being taken, and the remaining criteria in 10 CFR 2, App. C, to exercise discretion were satisfied.

A. 10 CFR 36.23 (a) requires in part that panoramic irradiator must have a sign bearing the radiation symbol "High radiation area", the sign may be removed, covered, or otherwise made ineffectiveness when sources are fully shielded.

B. At the time of the inspection the licensee informed the inspector that they will order signs for the door to the MCL Klorwic 70 irradiator.

C. _____

FE07
110