

SAFETY INSPECTION

1E07

RECEIVED  
NRC

1. LICENSEE

Bartlett Memorial Hospital  
Radiology Department  
3260 Hospital Drive  
Juneau, Alaska 99801

2. REGIONAL OFFICE

U.S. Nuclear Regulatory Commission  
1450 Maria Lane, Suite 210  
Walnut Creek, California 94596

3. DOCKET NUMBER(S)

4. LICENSE NUMBER(S)

5. DATE OF INSPECTION

50-17686-01

August 31, 1982

Licensee:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission's (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews, with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

1. Within the scope of this inspection, no violations were observed.

2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.

3. During this inspection certain of your activities, as checked below, were in violation of NRC requirements.

THIS IS A NOTICE OF VIOLATION which is required to be posted in accordance with 10 CFR 19.11.

A. \_\_\_\_\_ was not properly posted to indicate the presence of a \_\_\_\_\_ 10 CFR 20.203(b), (c), (d), (e) or 34.42.

B. Containers located in \_\_\_\_\_ were not properly labeled to indicate the presence of radioactive material. 10 CFR 20.203(f)(1), or (f)(2).

C. \_\_\_\_\_ of sealed sources were not performed at the proper frequencies. 10 CFR \_\_\_\_\_ License Condition Number \_\_\_\_\_

D. Records of \_\_\_\_\_ were not properly maintained. 10 CFR \_\_\_\_\_ or License Condition Number \_\_\_\_\_

E. Documents were not properly posted or otherwise made available. 10 CFR 19.11.

F. Reports or notifications of \_\_\_\_\_ were not made in accordance with 10 CFR \_\_\_\_\_ or License Condition Number \_\_\_\_\_

H. Contrary to statements made in the letter dated May 13, 1978, which is referenced in License Condition 14, elution, preparation and injection

I. areas are not surveyed each day that dose elution and preparation takes place.

J. \_\_\_\_\_

K. \_\_\_\_\_  
8302230118 820910  
NMS LIC30  
50-17686-01 PDR

I hereby state that within 30 days the actions described by me to the Inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

*James R. Bur*  
SIGNATURE - LICENSEE

9/20/82  
DATE

*B.A. Riedinger*  
SIGNATURE - NRC INSPECTOR

9/10/82  
DATE