

LAKEWOOD HOSPITAL

14519 DETROIT AVENUE  
LAKEWOOD 7, OHIO

ARTHUR A. LEPINOT, SUPERINTENDENT

February 16, 1965

Mr. Roy C. Hageman, Director  
Region III  
United States Atomic Energy Commission  
Division of Compliance  
Oakbrook Professional Building, Suite 410  
Oak Brook, Illinois

Dear Mr. Hageman:

In response to your letter of February 5, 1965, I wish to inform you of the constructive steps taken with regard to our Radio Isotope Unit, as follows:

- A. Effective February 15, 1965, we reduced our Byproduct material inventory from 10 microcuries of Cobalt 60, to the required 5 microcuries.
- B. Records of waste disposal are now being kept, in accordance with Regulation 10 CFR 20. 401 (b), effective February 1, 1965.
- C. Specific records of survey results of the Isotope Laboratory are being kept, as of February 1, 1965.

We appreciate very much having the survey by Mr. Ashley and are happy to comply with the regulations. If there is anything further that needs to be done, we will be very pleased to hear from you.

Very truly yours,

*Arthur A. Lepinot*  
Arthur A. Lepinot  
Superintendent

8209220052 820729  
PDR FOIA  
MIHAL82-310 PDR

AAL:del

FEB 18 1965