

MAY 24 1968

DML:IB:JEB (02455)

Lakewood Hospital
Department of Nuclear Medicine
14519 Detroit Avenue
Lakewood, Ohio 44107

Attention: Allen E. Walker, M.D.

Gentlemen:

In support of your request for possession and use of a molybdenum 99/technetium 99m generator, you should submit the following:

1. The manufacturer's name and model number of the survey meter which you will have available and which is capable of measuring exposure rates as high as one roentgen per hour.
2. Your method for evaluating the radiation exposures to hands and arms of personnel handling the generators.

Sincerely yours,

John E. Bowyer
Isotopes Branch
Division of Materials
Licensing

bcc: CO, Region III

8209220077 820729
PDR FOIA
MIHAL82-310

PDR

DML:IB

JEBowyer/vjh

5/24/68

MAY 31 1968