UNITED STATES NUCLEAR REGULATORY COMMISSION

REGIONIV

611 RYAN PLAZA DRIVE, SUITE 400 ARLINGTON, TEXAS 76011-8064 MAR 29 100A

Docket: 030-26841

License: 42-23539-01AF

CAL 4-94-01

Department of the Air Force USAF Radioisotope Committee ATTN: Lt. Colonel Joseph J. Donnelly HO AFMOA/SGPR 8901 18th Street Brooks AFB, Texas 78235-5217

SUBJECT. RESPONSE TO CONFIRMATORY ACTION LETTER

This is to acknowledge receipt of your letter dated March 8, 1994, in response to NRC's Confirmatory Action Letter (CAL) dated January 18, 1994. The CAL was in regard to the misadministration at Keesler Air Force Base (AFB). Mississippi.

The LAL confirmed with you the following actions:

- The Omnitron high dose rate afterloader at Keesler AFB would not be used for patient treatment until the device had been reviewed and serviced by the manufacturer, Omnitron, and
- 2. You would confirm, in writing to this office, the results of the above action prior to resuming use of the Omnitron.

Your response letter affirmed that the Omnitron high dose afterloader was serviced by the manufacturer on March 3, 1994, and was determined by the manufacturer to be fully operational and ready to be used for patient treatment. A written record of the inspection and servicing was attached. These actions, therefore, satisfy all the conditions of the CAL.

Should you have any questions regarding this letter please contact Ms. Linda McLean at (817) 860-8116.

Sincerely.

Collins Director Division of Radiation Safety and Safeguards

bcc: DMB - Original (IE-07) LJCallan SJCollins DDChamberlain LWCamper RAScarano WLFisher CLCain MLMcLean JASmith, NMSS (6 H3) CMHosey, RII NMIB MIS System RIV Files (2)

RIV:NMIB	C:NMIBLAC	D:DRSS	
MLMcLean Myn	CLCain	SJeolins	
03/28/94	03/28/94	03/20/94	

OMB - Original (IE-07) LJCallan SJCollins DDChamberlain LWCamper RAScarano WLFisher CLCain MLMcLean JASmith, NMSS (6 H3) CMHosey, RII NMIB MIS System RIV Files (2)

RIV:NMIB	C:NMIBLAC	D:DRSS
MLMcLean Myn	CLCain	SJeollins
03/28/94	03/28/94	03/28/794