

Inspect

DEC 02 1987

Caguas Sono-Nuclear and
Vascular Center
ATTN: Carmen Caballero, M.D.
President
P. O. Box 6960
Caguas, PR 00626

Gentlemen:

SUBJECT: CONFIRMATION OF ACTION (NRC LICENSE NO. 52-17273-01)

This refers to the discussions during a telephone conversation on November 25, 1987, between Mr. J. Potter of my staff and J. Scott-Cora, M.D., concerning the actions you will take regarding the NRC inspection conducted at your facility on November 18, 1987.

Based on matters discussed, it is our understanding that:

1. Technecium 99 DTPA aerosol will not be used until an amendment to your license is issued authorizing such use.
2. The dose calibrator will be immediately tested for accuracy, linearity, and geometrical variation.
3. Any elution from the Technicium 99 generator found to have in excess of 0.15 μ Ci of Mo-99 per mCi of Tc-99 will not be used for any medical procedures.

If your understanding of our discussion is different from that stated above, please inform this office promptly. Issuance of this letter does not preclude subsequent enforcement by Order.

Sincerely,
ORIGINAL SIGNED BY:
J. NELSON GRADE

J. Nelson Grace
Regional Administrator

CAL NO. 52-17273-01/87-01
CERTIFIED MAIL
RETURN RECEIPT REQUESTED

8712070234