AUBLIC 0 030-33408

MAR 2 4 1994

Port Huron Heart Center ATTN: Bashar Samman, M.D. Radiation Safety Officer 1201 Stone Street, Suite 1 Port Huron, MI 48060-3563 License No. 21-26548-01

Dear Dr. Samman:

This will acknowledge receipt of your letter dated February 19, 1994 indicating that the requirements of 10 CFR 35.32, "Quality Management Program" (QMP), do not apply to your program. Based on your declaration, we agree that a QMP need not be implemented at this time. However, in the future, should you wish to institute procedures in accordance with the provisions of Section 35.32(a)(1), it will be necessary to submit a QMP to our office for review and approval prior to its implementation.

Should you have any questions or need clarification regarding this matter, you may call us at (708) 829-9887.

Sincerely,

Original Signed By John R. Madera, Chief Materials Licensing Section

Frazier/jaw 03/11/94

Madera 03//4/94

9404060220 940324 PDR ADOCK 03033408 C PDR

20 0H

010062

PORT HURON HEART CENTER

CARDIOLOGY ASSOCIATES OF PORT HURON, P.C. 1201 Stone Street, Suite 1 Port Huron, MI 48060-3563 Office (810) 985-9681 \$\ \text{Fax} (810) 985-5310

21-26548-01 030-33408

JOHNSON K. ZACHARIAS, M.D. VRAJMOHAN C. PARIKH, M.D., F.A.C.C. SIVAJI GUNDLAPALLI, M.D., F.A.C.C.

VIRENDER D. PAREKH, M.D., F.A.C.C. MADHUSUDHAN C. REDDY, M.D., F.A.C.C. B. RAMESH REDDY, M.D., F.A.C.C. BASHAR SAMMAN, M.D.

February 19, 1994

Mr. Peter Lee Nuclear Regulatory Commission Region III, Materials Licensing Section 801 Warrenville Rd. Lisle, ILLINOIS 60532-4351

RE: Control #96198

Dear Sir:

We have reviewed regulatory guide 8.33 and have determined that it does not apply to our program since we do not perform any radiopharmaceutical therapy or use I-125 or I-131 in dosages of 30 uCi or above.

Thus, the quality management program does not apply to our facility. If there are any questions please call me.

Thank you.

Sincerely,

Bachar Samman, M.D. Radiation Safety Officer

BS:clc

RECEIVED REGION III